



Ref. Certif. No.

DE 3 - 31217

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Electronic measuring equipment

Name and address of the applicant

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19
74385 Pleidelsheim
GERMANY

Name and address of the manufacturer

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Name and address of the factory

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Ratings and principal characteristics

Rated voltage: 100-120 / 200-240 V
Rated frequency: 50/60 Hz
Rated power: 600 VA
Protection class: I
Air pressure: 6-8 bar

Trade mark (if any)

ZOLLER

Customer's Testing Facility (CTF) Stage used

CTF STAGE 1

Model/type Ref.

VENT450X-xxxxx, SMART450X-xxxxx
xxxxx: is used as serial number and can be 00050 to 99999

Additional information (if necessary)

Certificate DE 3 - 30927 issued 2019-05-08 is replaced by this version due to technical changes

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-1:2010/AMD1:2016

as shown in the Test Report Ref. No. which forms part of this certificate

028-713171903-000

This CB Test Certificate is issued by the National Certification Body

CB 095184 0023 Rev. 00

Date, 2020-04-03

(Abdul Sabbagh)



Product Service

Page 1 of 1

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany