



Ref. Certif. No.

DE 3 - 32064

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Electronic measuring equipment
Name and address of the applicant	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19 74385 Pleidelsheim GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	see next page
Trademark / Brand (if any)	ZOLLER
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	genius, threadCheck, edgeControl, 3dCheck, titan
Additional information (if necessary)	Certificate DE 3 – 31467 issued 2021-02-25 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713262145

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 01

Date, 2023-05-15

(Erich Thurner)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Ratings and principal characteristics (continued)

Rated voltage:	100-120 V and 200-240 V
Rated frequency:	50/60 Hz
Rated power:	600 VA
Protection class:	I
Overvoltage category:	II
Pollution degree:	2
Air pressure:	6-8 bar
Ambient:	5-35 °C, 10-90 % r.H (non-condensing)

Condition of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The detachable power cord set shall be complying with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard. The detachable power cord set shall not longer than 3 m.

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 01

Date, 2023-05-15



(Erich Thurner)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service