

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Electronic measuring equipment
Name and address of the applicant	<b>E. ZOLLER GmbH &amp; Co. KG</b> Gottlieb-Daimler-Str. 19 74385 Pleidelsheim GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	<b>see page 2</b>
Trade mark (if any)	ZOLLER
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	venturion 600, smartCheck 600, phoenix 600, millCheck, hobCheck, redomatic 600, reamCheck 600, scanTast, torquematic 600, redomatic 600 XL
Additional information (if necessary)	Certificate DE 3 – 31338 issued 2020-08-11 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713214033

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0037 Rev. 00

Date, 2021-10-13



( Hendrik Elze )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

Ratings and principal characteristics (continued)

**venturion 600, smartCheck 600, phoenix 600, millCheck, hobCheck, redomatic 600,  
reamCheck 600, scanTast, redomatic 600 XL:**

Rated voltage: 100-120 V and 200-240 V  
Rated frequency: 50/60 Hz  
Rated power: 600 VA  
Protection class: class I  
air pressure: 6-8 bar

**torquematic 600:**

Rated voltage: 120/240 V  
Rated frequency: 50/60 Hz  
Rated power: 720 VA  
Protection class: I  
air pressure: 6-8 bar

**Conditions of acceptability:**

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The detachable power cord set shall be complying with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard.
- The cockpit shall be fixed mounted with building structure.

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0037 Rev. 00

Date, 2021-10-13



( Hendrik Elze )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service