



America

CERTIFICATE

No. U8 095184 0020 Rev. 00

Holder of Certificate: **E. ZOLLER GmbH & Co. KG**
Gottlieb-Daimler-Str. 19
74385 Pleidelsheim
GERMANY

Production Facility(ies): 095184

Certification Mark:



Product: **Electronic measuring equipment**

Model(s): **VENT450X-xxxxx, SMART450X-xxxxx**
xxxx: is used as serial number and can be 00050 to 99999

Parameters:

Rated voltage:	100-120 / 200-240 V
Rated frequency:	50/60 Hz
Rated power:	600 VA
Protection class:	I
Air pressure:	6-8 bar

Tested according to: UL 61010-1:2012/R:2018-11
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 028-713171903-000

Date, 2020-04-03

(Benedikt Pulver)