



Ref. Certif. No.

DE 3 - 30927

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**

Product

Electronic measuring equipment

Name and address of the applicant

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19
74385 Pleidelsheim
GERMANY

Name and address of the manufacturer

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Name and address of the factory

E. ZOLLER GmbH & Co. KG
Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Ratings and principal characteristics

Rated voltage: 100-120 / 200-240 V
Rated frequency: 50/60 Hz
Rated power: 600 VA
Protection class: I
Air pressure: 6-8 barHint:
Environmental conditions: 5-35 °C, 5-90 % r.H.
Lamp classification: exempt group

Trade mark (if any)

ZOLLER

Customer's Testing Facility (CTF) Stage used

CTF STAGE 1

Model/type Ref.

VENT450X-xxxxx, SMART450X-xxxxx
xxxxx: is used as serial number and can be 00050 to 99999

Additional information (if necessary)

Certificate DE 3 - 30570 issued 2017-11-14 is replaced by this version due to technical changes

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010

as shown in the Test Report Ref. No. which forms part of this certificate

028-713150112-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

CB 095184 0017 Rev. 00

Date, 2019-05-08

(Erich Thurner)



Product Service

Page 1 of 1

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany