

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product	Electronic measuring equipment
Name and address of the applicant	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19 74385 Pleidelsheim, GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	Rated voltage: 100-120 / 200-240 V Rated frequency: 50/60 Hz Rated power: 600 VA Protection class: I Air pressure: 6-8 bar Environmental conditions: 5-35 °C, 5-90 % r.H.
Trade mark (if any)	ZOLLER
Customer's Testing Facility (CTF) Stage used	CTF Stage 1
Model/type Ref.	VENT450X-xxxxx, SMART450X-xxxxx xxxxx: is used as serial number and can be 00050 to 99999
Additional information (if necessary)	Certificate DE 3 - 30420 issued 2017-04-13 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010
as shown in the Test Report Ref. No. which forms part of this certificate	028-713109524-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-11-14
CB 17 11 95184 013



Ralph Fischer



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service